



Biological Sterilizer Monitoring Program

Test Results

Week 9 (02/26/2018) - Week 12 (03/25/2018)

Account Id: C-152264
Customer: Piercing Experience - Atlanta

1654 McLendon Ave Ne
Atlanta, GA 30307-2153

Sterilizer	Test #	Result Date	Result	Result Description
#01 STEAM AUTOCLAVE: SCICAN, STATIM 2000	17871862	3/21/2018	Pass	No growth of test spores, growth of control spores
	17871863	3/12/2018	Pass	No growth of test spores, growth of control spores
	17871864	3/5/2018	Pass	No growth of test spores, growth of control spores
	17871865	3/2/2018	Pass	No growth of test spores, growth of control spores
	17871866	3/2/2018	Pass	No growth of test spores, growth of control spores

Data displayed in this report is based on information provided to Mesa Labs by the client(s) at the time the tests are processed.