



Biological Sterilizer Monitoring Program

Test Results

Week 14 (04/02/2018) - Week 17 (04/29/2018)

Account Id: C-152264
Customer: Piercing Experience - Atlanta

1654 McLendon Ave Ne
Atlanta, GA 30307-2153

Sterilizer	Test #	Result Date	Result	Result Description
#01 STEAM AUTOCLAVE: SCICAN, STATIM 2000	17871857	4/23/2018	Pass	No growth of test spores, growth of control spores
	17871858	4/19/2018	Pass	No growth of test spores, growth of control spores
	17871859	4/9/2018	Pass	No growth of test spores, growth of control spores
	17871860	4/4/2018	Pass	No growth of test spores, growth of control spores

Data displayed in this report is based on information provided to Mesa Labs by the client(s) at the time the tests are processed.