



Biological Sterilizer Monitoring Program

Test Results

Week 18 (04/30/2018) - Week 21 (05/27/2018)

Account Id: C-152264
Customer: Piercing Experience - Atlanta

1654 McLendon Ave Ne
Atlanta, GA 30307-2153

Sterilizer	Test #	Result Date	Result	Result Description
#01 STEAM AUTOCLAVE: SCICAN, STATIM 2000	17871853	5/24/2018	Pass	No growth of test spores, growth of control spores
	17871854	5/16/2018	Pass	No growth of test spores, growth of control spores
	17871855	5/13/2018	Pass	No growth of test spores, growth of control spores
	17871856	5/3/2018	Pass	No growth of test spores, growth of control spores

Data displayed in this report is based on information provided to Mesa Labs by the client(s) at the time the tests are processed.